

Quality Assessment Scheme for Acetylcholine receptor autoantibodies

The results from all participants can be found in the outstanding table, listed according to their Lab Number.


Method	all	commercial kits				In house assay	
		IBL	RSR RIA	ELISA Euroimmun	RSR ELISA		other
Cut-off		>0.4	≥0.5	≥0.5	≥0.45	≥0.5	none
Number of lab.		(nmol/L)	(nmol/L)	(nmol/L)	(nmol/L)	(nmol/L)	(nmol/L)
131	0,34	0,34					
199	0,41	0,41					
207	1,22					1,22	
242	0,88		0,88				
364	0,36	0,36					
379	0,40	0,40					
381	0,41	0,41					
382	0,60						0,60
385	0,80		0,80				
389	0,38	0,38					
393	0,33	0,33					
397	0,32	0,32					
398	0,35	0,35					
399	0,41	0,41					
401	2,28				2,28		
402	0,33	0,33					
403	0,41	0,41					
406	0,32	0,32					
414	0,40	0,40					
418	0,40		0,40				
419	0,37	0,37					
421	0,36					0,36	
423	0,70					0,70	
425	0,37	0,37					
426	0,38	0,38					
428	1,40				1,4		
500	0,37	0,37					

Statistical Results Sample Number 24-1A:

Method	all	commercial kits					In house assay
		IBL	RSR RIA	ELISA Euroimmun	RSR ELISA	other	
Mean	0,57	0,37	0,69	-	1,84	0,76	0,60
SD	0,44	0,03	0,26	-	-	0,43	-
Mean - 2SD	<0	0,30	0,18	-	-	<0	-
Mean + 2SD	1,44	0,43	1,21	-	-	1,63	-
Amount	27	18	3	0	2	3	1


Samples have been determined by the single collectives as described below:

IBL
 Negative < 0.25 nmol/L
 Positive > 0.40 nmol/L
 Equivocal 0.25 – 0.40 nmol/L



*The graphic is for information purposes only and not validated. Please read your results from the table.

Method	all	commercial kits					In house assay
		IBL	RSR RIA	ELISA Euroimmun	RSR ELISA	other	
Negative	1	1	0	0	0	0	0
Equivocal	16	13	1	0	1	1	0
Positive	10	4	2	0	1	2	1
Amount	27	18	3	0	2	3	1



*The graphic is for information purposes only and not validated. Please read your results from the table.

The results from all participants can be found in the outstanding table, listed according to their Lab Number.


Method	all	commercial kits					In house assay
		IBL	RSR RIA	ELISA Euroimmun	RSR ELISA	other	
Cut-off		>0.4	≥0.5	≥0.5	≥0.45	≥0.5	none
Number of lab.		(nmol/L)	(nmol/L)	(nmol/L)	(nmol/L)	(nmol/L)	(nmol/L)
131	0,07	0,07					
199	0,13	0,13					
207	0,09					0,09	
242	0,25		0,25				
364	0,25	0,25					
379	0,07	0,07					
381	0,2	0,20					
382	0						0,00
385	0,1		0,10				
389	0	0,00					
393	0	0,00					
397	0,2	0,20					
398	0,25	0,25					
399	0,1	0,10					
401	0,45				0,45		
402	0,03	0,03					
403	0,2	0,20					
406	0,25	0,25					
414	0,1	0,10					
418	0,1		0,10				
419	0	0					
421	0,1					0,10	
423	0					0,00	
425	0,09	0,09					
426	0,2	0,2					
428	0,376				0,376		
500	0,01	0,01					

Statistical Results Sample Number 24-1B:

Method	all	commercial kits					In house assay
		IBL	RSR RIA	ELISA Euroimmun	RSR ELISA	other	
Mean	0,13	0,12	0,15	-	0,41	0,06	0,00
SD	0,12	0,09	0,09	-	-	0,06	-
Mean - 2SD	<0	<0	<0	-	-	<0	-
Mean + 2SD	0,37	0,31	0,32	-	-	0,17	-
Amount	27	18	3	0	2	3	1


Samples have been determined by the single collectives as described below:

Negative < 0.25 nmol/L
 Positive > 0.40 nmol/L
 Equivocal 0.25 – 0.40 nmol/L



*The graphic is for information purposes only and not validated. Please read your results from the table.

Method	all	commercial kits					In house assay
		IBL	RSR RIA	ELISA Euroimmun	RSR ELISA	other	
Negative	27	18	3	0	2	3	1
Equivocal	0	0	0	0	0	0	0
Positive	0	0	0	0	0	0	0
Amount	27	18	3	0	2	3	1



*The graphic is for information purposes only and not validated. Please read your results from the table.

The results from all participants can be found in the outstanding table, listed according to their Lab Number.


Method	all	commercial kits					In house assay
		IBL	RSR RIA	ELISA Euroimmun	RSR ELISA	other	
Cut-off		>0.4	≥0.5	≥0.5	≥0.45	≥0.5	none
Number of lab.		(nmol/L)	(nmol/L)	(nmol/L)	(nmol/L)	(nmol/L)	(nmol/L)
131	1,19	1,19					
199	1,09	1,09					
207	4,72					4,72	
242	2,72		2,72				
364	1,05	1,05					
379	1,10	1,10					
381	0,92	0,92					
382	2,10						2,10
385	2,40		2,40				
389	1,31	1,31					
393	1,08	1,08					
397	1,14	1,14					
398	1,10	1,10					
399	1,25	1,25					
401	7,53				7,53		
402	1,09	1,09					
403	1,30	1,30					
406	1,20	1,20					
414	1,27	1,27					
418	2,40		2,40				
419	1,04	1,04					
421	2,28					2,28	
423	2,50					2,50	
425	1,16	1,16					
426	1,17	1,17					
428	7,20				7,20		
500	1,14	1,14					

Statistical Results Sample Number 24-1C:

Method	all	commercial kits					In house assay
		IBL	RSR RIA	ELISA Euroimmun	RSR ELISA	other	
Mean	2,02	1,14	2,51	-	7,37	3,17	2,10
SD	1,75	0,10	0,18	-	-	1,35	-
Mean - 2SD	<0	0,95	2,14	-	-	0,47	-
Mean + 2SD	5,51	1,34	2,88	-	-	5,87	-
Amount	27	18	3	0	2	3	1


Samples have been determined by the single collectives as described below:

Negative < 0.25 nmol/L
 Positive > 0.40 nmol/L
 Equivocal 0.25 – 0.40 nmol/L



*The graphic is for information purposes only and not validated. Please read your results from the table.

Method	all	commercial kits					In house assay
		IBL	RSR RIA	ELISA Euroimmun	RSR ELISA	other	
Negative	0	0	0	0	0	0	0
Equivocal	0	0	0	0	0	0	0
Positive	27	18	3	0	2	3	1
Amount	27	18	3	0	2	3	1



*The graphic is for information purposes only and not validated. Please read your results from the table.