

Quality Assessment Scheme for Salivary Steroids

Cortisol Sample Number 24-11

Your result: The results from all participants can be found in the outstanding table.

Method	LUM	ELISA						MS	Others
Supplier	all	IBL	IBL	Salimetrics	Demeditec	DRG	Others	In-House	Roche
Lab. Code	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL
1318	0,210					0,210			
3013	0,161	0,161							
3014	0,199							0,199	
3019	0,167							0,167	
3025	0,195							0,195	
3027	0,187							0,187	
3028	0,292								0,292
3030	0,174							0,174	
3031	0,271								0,271
3032	0,149							0,149	
3035	0,140		0,140						
3036	0,243								0,243
3037	0,120							0,120	
3040	0,174		0,174						
3041	0,198							0,198	
3047	0,188							0,188	
3048	0,164		0,164						
3049	0,180							0,180	
3056	0,196							0,196	
3059	0,232								0,232
3063	0,227								0,227
13004	0,190	0,190							
13007	0,218							0,218	
15033	0,190							0,190	
22236	0,209			0,209					
30001	0,185							0,185	
30006	0,210			0,210					
40001	0,176	0,176							
40002	0,141		0,141						
40003-1	0,209							0,209	
40003-2	0,168	0,168							
40004	0,202								0,202
40014	0,185							0,185	
40015	0,484				0,484				
40017	0,710					0,710			
40018	0,156		0,156						
40020	0,253								0,253
40021	0,179		0,179						
40022	0,182							0,182	

Quality Assessment Scheme for Salivary Steroids

Cortisol Sample Number 24-11

Your result: The results from all participants can be found in the outstanding table.


Method	LUM		ELISA				MS	Others	
Supplier	all	IBL	IBL	Salimetrics	Demeditec	DRG	Others	In-House	Roche
Lab. Code	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL
40024	0,148		0,148						
40027	0,160		0,160						
40029	0,543				0,543				
40031	0,240								0,240
40036	0,119		0,119						
40038	0,535				0,535				
40041	0,430				0,430				
40043	0,225			0,225					
40047	0,175	0,175							
40049	0,190		0,190						
40054	0,226							0,226	
40055	0,176							0,176	
40056	0,330						0,330		
50002	0,141		0,141						
60003	0,156		0,156						
70001	0,206		0,206						

Cortisol Sample Number 24-11

Statistical Results

Method	LUM		ELISA				MS	Others	
Supplier	all	IBL	IBL	Salimetrics	Demeditec	DRG	Others	In-House	Roche
Mean	0,224	0,174	0,160	0,215	0,498	0,460	0,330	0,185	0,245
SD	0,112	0,011	0,023	0,009	0,052	0,354		0,024	0,028
Mean - 2SD	0,000	0,152	0,113	0,197	0,393	0,000		0,138	0,190
Mean + 2SD	0,448	0,196	0,206	0,233	0,603	1,167		0,233	0,300
Amount	55	5	13	3	4	2	1	19	8

*The graphic is for information purposes only and not validated. Please read your results from the table.



Quality Assessment Scheme for Salivary Steroids

Cortisol Sample Number 24-12

Your result: The results from all participants can be found in the outstanding table.

Method		LUM	ELISA					MS	Others
Supplier	all	IBL	IBL	Salimetrics	Demeditec	DRG	Others	In-House	Roche
Lab. Code	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL
1318	0,062					0,062			
3013	0,048	0,048							
3014	0,062							0,062	
3019	0,040							0,040	
3025	0,069							0,069	
3027	0,068							0,068	
3028	0,118								0,118
3030	0,058							0,058	
3031	0,130								0,130
3032	0,051							0,051	
3035	0,060		0,060						
3036	0,004								0,004
3037	0,036							0,036	
3040	0,076		0,076						
3041	0,024							0,024	
3047	0,060							0,060	
3048	0,055		0,055						
3049	0,069							0,069	
3056	0,076							0,076	
3059	0,080								0,080
3063	0,075								0,075
13004	0,070	0,070							
13007	0,060							0,060	
15033	0,070							0,070	
22236	0,079			0,079					
30001	0,064							0,064	
30006	0,080			0,080					
40001	0,067	0,067							
40002	0,071		0,071						
40003-1	0,075							0,075	
40003-2	0,067	0,067							
40004	0,083								0,083
40014	0,060							0,060	
40015	0,235				0,235				
40017	0,380					0,380			
40018	0,047		0,047						
40020	0,108								0,108
40021	0,058		0,058						
40022	0,067							0,067	

Quality Assessment Scheme for Salivary Steroids

Cortisol Sample Number 24-12

Your result: The results from all participants can be found in the outstanding table.


Method		LUM	ELISA					MS	Others
Supplier	all	IBL	IBL	Salimetrics	Demeditec	DRG	Others	In-House	Roche
Lab. Code	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL
40024	0,054		0,054						
40027	0,070		0,070						
40029	0,225				0,225				
40031	0,082								0,082
40036	0,046		0,046						
40038	0,207				0,207				
40041	0,180				0,180				
40043	0,085			0,085					
40047	0,068	0,068							
40049	0,063		0,063						
40054	0,093							0,093	
40055	0,064							0,064	
40056	0,139						0,139		
50002	0,054		0,054						
60003	0,080		0,080						
70001	0,079		0,079						

Cortisol Sample Number 24-12

Statistical Results

Method		LUM	ELISA					MS	Others
Supplier	all	IBL	IBL	Salimetrics	Demeditec	DRG	Others	In-House	Roche
Mean	0,085	0,064	0,063	0,081	0,212	0,221	0,139	0,061	0,085
SD	0,060	0,009	0,012	0,003	0,024	0,225		0,015	0,038
Mean - 2SD	0,000	0,046	0,039	0,075	0,163	0,000		0,030	0,008
Mean + 2SD	0,204	0,082	0,086	0,088	0,260	0,671		0,092	0,162
Amount	55	5	13	3	4	2	1	19	8

*The graphic is for information purposes only and not validated. Please read your results from the table.



Quality Assessment Scheme for Salivary Steroids

Cortisol Sample Number 24-13

Your result: The results from all participants can be found in the outstanding table.

Method		LUM	ELISA					MS	Others
Supplier	all	IBL	IBL	Salimetrics	Demeditec	DRG	Others	In-House	Roche
Lab. Code	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL
1318	1,333					1,333			
3013	1,094	1,094							
3014	1,348							1,348	
3019	0,928							0,928	
3025	1,225							1,225	
3027	1,159							1,159	
3028	1,410								1,410
3030	1,004							1,004	
3031	1,446								1,446
3032	0,870							0,870	
3035	1,050		1,050						
3036	0,982								0,982
3037	0,797							0,797	
3040	1,025		1,025						
3041	1,270							1,270	
3047	1,001							1,001	
3048	1,319		1,319						
3049	1,023							1,023	
3056	1,130							1,130	
3059	1,322								1,322
3063	1,330								1,330
13004	1,320	1,320							
13007	1,230							1,230	
15033	1,110							1,110	
22236	1,098			1,098					
30001	1,141							1,141	
30006	1,340			1,340					
40001	1,166	1,166							
40002	0,891		0,891						
40003-1	1,157							1,157	
40003-2	1,033	1,033							
40004	1,005								1,005
40014	1,150							1,150	
40015	2,520				2,520				
40017	2,780					2,780			
40018	0,976		0,976						
40020	1,370								1,370
40021	1,162		1,162						
40022	1,017							1,017	

Quality Assessment Scheme for Salivary Steroids

Cortisol Sample Number 24-13

Your result: The results from all participants can be found in the outstanding table.

Method	LUM		ELISA				MS	Others	
Supplier	all	IBL	IBL	Salimetrics	Demeditec	DRG	Others	In-House	Roche
Lab. Code	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL	µg/dL
40024	1,060		1,060						
40027	1,070		1,070						
40029	2,729				2,729				
40031	1,310								1,310
40036	0,805		0,805						
40038	2,549				2,549				
40041	1,920				1,920				
40043	1,200			1,200					
40047	1,120	1,120							
40049	1,155		1,155						
40054	1,109							1,109	
40055	1,030							1,030	
40056	2,179						2,179		
50002	1,094		1,094						
60003	0,851		0,851						
70001	1,077		1,077						

Cortisol Sample Number 24-13

Statistical Results

Method	LUM		ELISA				MS	Others	
Supplier	all	IBL	IBL	Salimetrics	Demeditec	DRG	Others	In-House	Roche
Mean	1,269	1,147	1,041	1,213	2,430	2,057	2,179	1,089	1,272
SD	0,454	0,108	0,138	0,121	0,352	1,023		0,138	0,178
Mean - 2SD	0,361	0,930	0,765	0,970	1,725	0,010		0,813	0,916
Mean + 2SD	2,176	1,363	1,317	1,456	3,134	4,103		1,366	1,628
Amount	55	5	13	3	4	2	1	19	8

*The graphic is for information purposes only and not validated. Please read your results from the table.

