


Quality Control Certificate



PR3 IgG ELISA (c-ANCA), capture version

REF 30112111

LOT 26010075

 2027-06-30
YYYY-MM-DD

Final Release Results:

	Conc. IU/mL	Measured value OD	Acceptable range		
			lower limit OD	target value OD	upper limit OD
CAL 1	0	0.061			< 0.100
CAL 6	100	2.121	> 1.000		
CONTROL -	-	0.126			≤ 0.300
CONTROL +	-	1.314	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value IU/mL	Acceptable range		
		lower limit IU/mL	target value IU/mL	upper limit IU/mL
CONTROL -	0.4			1.3
CONTROL +	20	13	20	27

Factor for qualitative assay evaluation	0.22
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2026-02-10

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.:
E-MAIL:
WEB:

+ 49 (0) 40 532891 -0 Fax: -11
ibl@tecan.com
www.tecan.com/ibl