

Quality Control Certificate



PR3 IgG ELISA (c-ANCA), capture version

REF 30112111

LOT 26080094

2027-11-30
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	IU/mL	OD	lower limit OD	target value OD	upper limit OD
CAL 1	0	0.044			< 0.100
CAL 6	100	2.168	> 1.000		
CONTROL -	-	0.164			≤ 0.300
CONTROL +	-	1.501	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
	IU/mL	lower limit IU/mL	target value IU/mL	upper limit IU/mL
CONTROL -	0.8			1.3
CONTROL +	25	13	20	27

Factor for qualitative assay evaluation	0.22
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-08-31

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.:
E-MAIL:
WEB:

+ 49 (0) 40 532891 -0 Fax: -11
ibl@tecan.com
www.tecan.com/ibl