

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Adenovirus IgG**
 Prod.No.: **30113466**
 Lot: **ADVG-157**
 Expiry: **2027-03-31**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0010-157	2025-09	2027-03-31
CONJ	1	KG0010-157	2025-09	2027-03-31
CONTROL -	1	NG0010-157	2025-09	2027-03-31
CUT OFF	1	CG0010-157	2025-09	2027-03-31
CONTROL +	1	PG0010-157	2025-09	2027-03-31
WASH BUF 20x	1	W-1119	2025-08	2027-08-31
SOLN STOP	1	S-1050	2025-08	2027-08-31
SUB TMB	1	T-078	2025-04	2028-04-28
DIL	1	D-494	2025-06	2027-06-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.011	< 0.200 < Cut-off
CUT OFF	0.582	0.150-1.300
CONTROL +	1.283	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	24.30	13.37 - 35.24

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-10-09

Approved by: 
 Quality Control

Approved on: 2025-10-09



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany