



**GOLD
STANDARD
DIAGNOSTICS**



Certificate of Analysis (CoA)


Product: **Adenovirus IgM**
 Prod.No.: **30113467**
 Lot: **ADVM-156**
 Expiry: **2027-01-31**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0010-156	2025-07	2027-01-31
CONJ	1	KM0010-156	2025-07	2027-01-31
CONTROL -	1	NM0010-156	2025-07	2027-01-31
CUT OFF	1	CM0010-156	2025-07	2027-01-31
CONTROL +	1	PM0010-156	2025-07	2027-01-31
WASH BUF 20x	1	W-1116	2025-06	2027-06-30
SOLN STOP	1	S-1047	2025-05	2027-05-31
SUB TMB	1	T-078	2025-04	2028-04-28
DIL M	1	DM-346	2025-06	2027-06-30


Component	OD Value	OD Validation Criteria
CONTROL -	0.022	< 0.200 < Cut-off
CUT OFF	0.481	0.150-1.300
CONTROL +	0.730	> Cut-off


Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	15.86	>10 - 22.99

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-08-12
 Approved by: 
 Quality Control

Approved on: 2025-08-12

 Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany

 IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany