

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Aspergillus fumigatus IgG**

Prod.No.: **30113468**

Lot: **ASPG-059NR**

Expiry: **2026-10-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0680-059	2025-04	2026-10-31
CONJ	1	KG0680-059	2025-04	2026-10-31
CONTROL -	1	NG0680-059	2025-04	2026-10-31
CUT OFF	1	CG0680-059	2025-04	2026-10-31
CONTROL +	1	PG0680-059	2025-04	2026-10-31
WASH BUF 20x	1	W-1108	2025-02	2027-02-28
SOLN STOP	1	S-1039	2025-01	2027-01-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-490	2024-12	2026-12-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.010	< 0.200 < Cut-off
CUT OFF	0.640	0.150-1.300
CONTROL +	1.204	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	20.36	11.20 - 29.52

Final Report issued: 2025-04-16

Approved by: 
Quality Control

Approved on: 2025-04-17



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany