

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Aspergillus fumigatus IgM**

Prod.No.: **30113471**

Lot: **ASPM-060**

Expiry: **2026-11-30**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0680-060	2025-05	2026-11-30
CONJ	1	KM0680-060	2025-05	2026-11-30
CONTROL -	1	NM0680-060	2025-05	2026-11-30
CUT OFF	1	CM0680-060	2025-05	2026-11-30
CONTROL +	1	PM0680-060	2025-05	2026-11-30
WASH BUF 20x	1	W-1111	2025-03	2027-03-31
SOLN STOP	1	S-1042	2025-02	2027-02-28
SUB TMB	1	T-077	2022-09	2027-03-31
DIL M	1	DM-344	2025-04	2027-04-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.018	< 0.200 < Cut-off
CUT OFF	0.632	0.150-1.300
CONTROL +	1.360	> Cut-off

Final Report issued: 2025-06-06

Approved by: Hoezalp
Quality Control

Approved on: 2025-06-06



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany