

**GOLD
STANDARD
DIAGNOSTICS**



Certificate of Analysis (CoA)

Product: **Aspergillus fumigatus IgM**
 Prod.No.: **30113471**
 Lot: **ASPM-065**
 Expiry: **2027-07-31**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0680-065	2026-01	2027-07-31
CONJ	1	KM0680-065	2026-01	2027-07-31
CONTROL -	1	NM0680-065	2026-01	2027-07-31
CUT OFF	1	CM0680-065	2026-01	2027-07-31
CONTROL +	1	PM0680-065	2026-01	2027-07-31
WASH BUF 20x	1	W-1132	2026-01	2028-01-31
SOLN STOP	1	S-1066	2025-12	2027-12-31
SUB TMB	1	T-079	2025-12	2028-12-10
DIL M	1	DM-351	2026-01	2028-01-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.005	< 0.200 < Cut-off
CUT OFF	0.610	0.150-1.300
CONTROL +	1.286	> Cut-off


Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	20.72	11.39 - 30.04


The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-02-12

Approved by: Hoezalp
 Quality Control

Approved on: 2026-02-12

 Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany

 IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany