

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Aspergillus fumigatus IgM**

Prod.No.: **30113471**

Lot: **ASPM-066**

Expiry: **2027-11-30**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0680-066	2026-05	2027-11-30
CONJ	1	KM0680-066	2026-05	2027-11-30
CONTROL -	1	NM0680-066	2026-05	2027-11-30
CUT OFF	1	CM0680-066	2026-05	2027-11-30
CONTROL +	1	PM0680-066	2026-05	2027-11-30
WASH BUF 20x	1	W-1141	2026-04	2028-04-30
SOLN STOP	1	S-1072	2026-03	2028-03-31
SUB TMB	1	T-080	2026-02	2029-02-23
DIL M	1	DM-353	2026-03	2028-03-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.003	< 0.200 < Cut-off
CUT OFF	0.531	0.150-1.300
CONTROL +	0.982	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	19.02	10.46 - 27.58

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-05-18

Approved by: 
Quality Control

Approved on: 2026-05-19



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany