

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Brucella IgG**
 Prod.No.: **30113479**
 Lot: **BRUG-132**
 Expiry: **2027-07-31**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0050-132	2026-01	2027-07-31
CONJ	1	KG0050-132	2026-01	2027-07-31
CONTROL -	1	NG0050-132	2026-01	2027-07-31
CUT OFF	1	CG0050-132	2026-01	2027-07-31
CONTROL +	1	PG0050-132	2026-01	2027-07-31
WASH BUF 20x	1	W-1131	2026-01	2028-01-31
SOLN STOP	1	S-1065	2025-12	2027-12-31
SUB TMB	1	T-079	2025-12	2028-12-10
DIL	1	D-498	2025-12	2027-12-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.004	< 0.200 < Cut-off
CUT OFF	0.587	0.150-1.300
CONTROL +	0.972	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	16.16	>10 - 23.43

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-01-26

Approved by: 
 Quality Control

Approved on: 2026-01-26



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany