


Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Brucella IgG**

Prod.No.: **30113479**

Lot: **BRUG-134**

Expiry: **2027-08-31**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0050-134	2026-02	2027-08-31
CONJ	1	KG0050-134	2026-02	2027-08-31
CONTROL -	1	NG0050-134	2026-02	2027-08-31
CUT OFF	1	CG0050-134	2026-02	2027-08-31
CONTROL +	1	PG0050-134	2026-02	2027-08-31
WASH BUF 20x	1	W-1136	2026-02	2028-02-29
SOLN STOP	1	S-1068	2026-01	2028-01-31
SUB TMB	1	T-079	2025-12	2028-12-10
DIL	1	D-499	2026-01	2028-01-31


Component	OD Value	OD Validation Criteria
CONTROL -	0.010	< 0.200 < Cut-off
CUT OFF	0.582	0.150-1.300
CONTROL +	0.947	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	20.07	11.04 - 29.10

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-03-11

Approved by: 
Quality Control

Approved on: 2026-03-11



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany