



**GOLD  
STANDARD  
DIAGNOSTICS**



# Certificate of Analysis (CoA)

Product: **Candida albicans IgG**  
 Prod.No.: **30113618**  
 Lot: **CANG-209**  
 Expiry: **2027-04-30**  
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
<b>MTP</b>	1	MG0060-209	2025-10	2027-04-30
<b>CONJ</b>	1	KG0060-209	2025-10	2027-04-30
<b>CONTROL -</b>	1	NG0060-209	2025-10	2027-04-30
<b>CUT OFF</b>	1	CG0060-209	2025-10	2027-04-30
<b>CONTROL +</b>	1	PG0060-209	2025-10	2027-04-30
<b>WASH BUF 20x</b>	1	W-1125	2025-10	2027-10-31
<b>SOLN STOP</b>	1	S-1056	2025-09	2027-09-30
<b>SUB TMB</b>	1	T-078	2025-04	2028-04-28
<b>DIL</b>	1	D-495	2025-09	2027-09-30

Component	OD Value	OD Validation Criteria
<b>CONTROL -</b>	0.007	< 0.200 < Cut-off
<b>CUT OFF</b>	0.568	0.150-1.300
<b>CONTROL +</b>	1.192	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
<b>CONTROL +</b>	23.14	12.73 - 33.55

The target value and target range of **CONTROL +** may be used optionally.  
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-11-17

Approved by:   
 Quality Control

Approved on: 2025-11-17



Gold Standard Diagnostics Frankfurt GmbH  
 Waldstrasse 23 A6  
 63128 Dietzenbach, Germany



IBL International GmbH  
 Flughafenstrasse 52a, 22335 Hamburg  
 Germany