


Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Candida albicans IgM**

Prod.No.: **30113619**

Lot: **CANM-211**

Expiry: **2027-09-30**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0060-211	2026-03	2027-09-30
CONJ	1	KM0060-211	2026-03	2027-09-30
CONTROL -	1	NM0060-211	2026-03	2027-09-30
CUT OFF	1	CM0060-211	2026-03	2027-09-30
CONTROL +	1	PM0060-211	2026-03	2027-09-30
WASH BUF 20x	1	W-1137	2026-02	2028-02-29
SOLN STOP	1	S-1069	2026-02	2028-02-29
SUB TMB	1	T-079	2025-12	2028-12-10
DIL M	1	DM-352	2026-03	2028-02-29

Component	OD Value	OD Validation Criteria
CONTROL -	0.007	< 0.200 < Cut-off
CUT OFF	0.610	0.150-1.300
CONTROL +	1.419	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	24.01	13.21 - 34.81

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2026-03-19

Approved by: _____

Quality Control

Approved on: 2026-03-19



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany