


Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Hantavirus IgG**
 Prod.No.: **30114042**
 Lot: **HANG-045N**
 Expiry: **2026-05-31**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0670-045	2024-11	2026-05-31
CONJ	1	KG0670-045	2024-11	2026-05-31
CONTROL -	1	NG0670-045	2024-11	2026-05-31
CUT OFF	1	CG0670-045	2024-11	2026-05-31
CONTROL +	1	PG0670-045	2024-11	2026-05-31
WASH BUF 20x	1	W-1100	2024-10	2026-10-31
SOLN STOP	1	S-1031	2024-09	2026-09-30
SUB TMB	1	T-076	2022-05	2026-11-30
DIL	1	D-488	2024-08	2026-08-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.018	< 0.200 < Cut-off
CUT OFF	0.704	0.150-1.300
CONTROL +	1.520	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	22.83	12.56 - 33.10

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2024-12-05

Approved by: ACc
 Quality Control

Approved on: 2024-12-05



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany