


Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Influenza Virus B IgM**

Prod.No.: **30114054**

Lot: **INFBM-216**

Expiry: **2027-04-30**

IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0300-216	2025-10	2027-04-30
CONJ	1	KM0300-216	2025-10	2027-04-30
CONTROL -	1	NM0300-216	2025-10	2027-04-30
CUT OFF	1	CM0300-216	2025-10	2027-04-30
CONTROL +	1	PM0300-216	2025-10	2027-04-30
WASH BUF 20x	1	W-1125	2025-10	2027-10-31
SOLN STOP	1	S-1056	2025-09	2027-09-30
SUB TMB	1	T-078	2025-04	2028-04-28
DIL M	1	DM-348	2025-10	2027-10-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.004	< 0.200 < Cut-off
CUT OFF	0.582	0.150-1.300
CONTROL +	1.163	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	22.26	12.24 - 32.27

The target value and target range of **CONTROL +** may be used optionally.

The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-11-10

Approved by: Hoezalp
Quality Control

Approved on: 2025-11-10



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany