

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Measles Virus IgG**
 Prod.No.: **30114056**
 Lot: **MEAG-146**
 Expiry: **2027-04-30**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0330-146	2025-10	2027-04-30
CONJ	1	KG0330-146	2025-10	2027-04-30
CONTROL -	1	NG0330-146	2025-10	2027-04-30
CUT OFF	1	CG0330-146	2025-10	2027-04-30
CONTROL +	1	PG0330-146	2025-10	2027-04-30
WASH BUF 20x	1	W-1126	2025-11	2027-11-30
SOLN STOP	1	S-1057	2025-09	2027-09-30
SUB TMB	1	T-078	2025-04	2028-04-28
DIL	1	D-495	2025-09	2027-09-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.014	< 0.200 < Cut-off
CUT OFF	0.776	0.150-1.300
CONTROL +	1.638	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	22.17	12.19 - 32.15

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-11-17

Approved by: 
 Quality Control

Approved on: 2025-11-17



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany