Certificate of Analysis (CoA)

Product:	**Measles Virus IgG**
Prod.No.:	**30114056**
Lot:	**MEAG-150**
Expiry:	**2028-02-29**
IFU:	**Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0330-150	2026-08	2028-02-29
CONJ	1	KG0330-150	2026-08	2028-02-29
CONTROL -	1	NG0330-150	2026-08	2028-02-29
CUT OFF	1	CG0330-150	2026-08	2028-02-29
CONTROL +	1	PG0330-150	2026-08	2028-02-29
WASH BUF 20x	1	W-1149	2026-08	2028-08-31
SOLN STOP	1	S-1080	2026-07	2028-07-31
SUB TMB	1	T-080	2026-02	2029-02-23
DIL	1	D-503	2026-06	2028-06-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.012	< 0.200 < Cut-off
CUT OFF	0.582	0.150-1.300
CONTROL +	1.395	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	26.26	14.44 - 38.07

The target value and target range of CONTROL + may be used optionally.

The positive control is valid if the absorbance value of CONTROL + > CUT OFF.

Final Report issued: 2026-08-31

Approved by: _____

Quality Control

Approved on: 2026-08-31

Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany

IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany