

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Respiratory Syncytial Virus IgG**

Prod.No.: **30114070**

Lot: **RSVG-177NR**

Expiry: **2026-09-30**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG0380-177	2025-03	2026-09-30
CONJ	1	KG0380-177	2025-03	2026-09-30
CONTROL -	1	NG0380-177	2025-03	2026-09-30
CUT OFF	1	CG0380-177	2025-03	2026-09-30
CONTROL +	1	PG0380-177	2025-03	2026-09-30
WASH BUF 20x	1	W-1107	2025-01	2027-01-31
SOLN STOP	1	S-1038	2024-12	2026-12-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-490	2024-12	2026-12-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.007	< 0.200 < Cut-off
CUT OFF	0.608	0.150-1.300
CONTROL +	1.143	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	22.08	12.14 - 32.01

Final Report issued: 2025-04-01

Approved by: *S. Herrenschmidt*
Quality Control

Approved on: 2025-04-01



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany