



**GOLD
STANDARD
DIAGNOSTICS**



Certificate of Analysis (CoA)

Product: **Respiratory Syncytial Virus IgM**
 Prod.No.: **30114071**
 Lot: **RSVM-178**
 Expiry: **2026-11-30**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0380-178	2025-05	2026-11-30
CONJ	1	KM0380-178	2025-05	2026-11-30
CONTROL -	1	NM0380-178	2025-05	2026-11-30
CUT OFF	1	CM0380-178	2025-05	2026-11-30
CONTROL +	1	PM0380-178	2025-05	2026-11-30
WASH BUF 20x	1	W-1111	2025-03	2027-03-31
SOLN STOP	1	S-1042	2025-02	2027-02-28
SUB TMB	1	T-077	2022-09	2027-03-31
DIL M	1	DM-344	2025-04	2027-04-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.011	< 0.200 < Cut-off
CUT OFF	0.500	0.150-1.300
CONTROL +	1.100	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	23.07	12.69 - 33.45

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-06-02

Approved by: 
 Quality Control

Approved on: 2025-06-02



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany