

Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Clostridium tetani toxin 5S IgG plus**
 Prod.No.: **30114074**
 Lot: **PTETG-153**
 Expiry: **2027-04-30**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MG5043-153	2025-10	2027-04-30
CONJ	1	KG5043-153	2025-10	2027-04-30
CAL A	1	SAG5043-153	2025-10	2027-04-30
CAL B	1	SBG5043-153	2025-10	2027-04-30
CAL C	1	SCG5043-153	2025-10	2027-04-30
CAL D	1	SDG5043-153	2025-10	2027-04-30
CAL E	1	SEG5043-153	2025-10	2027-04-30
CONTROL	1	CG5043-153	2025-10	2027-04-30
WASH BUF 20x	1	W-1126	2025-11	2027-11-30
SOLN STOP	1	S-1057	2025-09	2027-09-30
SUB TMB	1	T-078	2025-04	2028-04-28
DIL	1	D-495	2025-09	2027-09-30

Component	Concentration	OD Value	OD Validation Criteria
CAL A	0.0 IU/mL	0.013	< 0.200
CAL B	0.1 IU/mL	0.331	> 0.050 & > Standard A
CAL C	0.5 IU/mL	0.825	> Standard B
CAL D	1.0 IU/mL	1.254	> Standard C
CAL E	5.0 IU/mL	2.122	> 0.800 & > Standard D

Component	Target Value (IU/mL)	Target Range (IU/mL)
CONTROL	2.14	1.17 - 3.10

Final Report issued: 2025-11-17

Approved by: 
 Quality Control

Approved on: 2025-11-17



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany