


Certificate of Analysis (CoA)



**GOLD
STANDARD
DIAGNOSTICS**



Product: **Varicella-Zoster Virus (VZV) IgA**

Prod.No.: **30114080**

Lot: **VZVA-187NR**

Expiry: **2026-08-31**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MA0490-187	2025-02	2026-08-31
CONJ	1	KA0490-187	2025-02	2026-08-31
CONTROL -	1	NA0490-187	2025-02	2026-08-31
CUT OFF	1	CA0490-187	2025-02	2026-08-31
CONTROL +	1	PA0490-187	2025-02	2026-08-31
WASH BUF 20x	1	W-1104	2024-12	2026-12-31
SOLN STOP	1	S-1035	2024-10	2026-10-31
SUB TMB	1	T-077	2022-09	2027-03-31
DIL	1	D-489	2024-10	2026-10-31

Component	OD Value	OD Validation Criteria
CONTROL -	0.014	< 0.200 < Cut-off
CUT OFF	0.602	0.150-1.300
CONTROL +	1.367	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	23.91	13.15 - 34.66

Final Report issued: 2025-02-24

Approved by: HOe
Quality Control

Approved on: 2025-02-24



Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany



IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany