



**GOLD
STANDARD
DIAGNOSTICS**



Certificate of Analysis (CoA)

Product: **Varicella-Zoster Virus (VZV) IgM**
 Prod.No.: **30114082**
 Lot: **VZVM-189**
 Expiry: **2026-10-31**
 IFU: **Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0490-189	2025-04	2026-10-31
CONJ	1	KM0490-189	2025-04	2026-10-31
CONTROL -	1	NM0490-189	2025-04	2026-10-31
CUT OFF	1	CM0490-189	2025-04	2026-10-31
CONTROL +	1	PM0490-189	2025-04	2026-10-31
WASH BUF 20x	1	W-1109	2025-02	2027-02-28
SOLN STOP	1	S-1040	2025-02	2027-02-28
SUB TMB	1	T-077	2022-09	2027-03-31
DIL M	1	DM-344	2025-04	2027-04-30

Component	OD Value	OD Validation Criteria
CONTROL -	0.018	< 0.200 < Cut-off
CUT OFF	0.564	0.150-1.300
CONTROL +	0.900	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	18.33	10.08 - 26.58

The target value and target range of **CONTROL +** may be used optionally.
 The positive control is valid if the absorbance value of **CONTROL +** > **CUT OFF**.

Final Report issued: 2025-05-08

Approved by: 
 Quality Control

Approved on: 2025-05-12



Gold Standard Diagnostics Frankfurt GmbH
 Waldstrasse 23 A6
 63128 Dietzenbach, Germany



IBL International GmbH
 Flughafenstrasse 52a, 22335 Hamburg
 Germany