Certificate of Analysis (CoA)

Product:	**Varicella-Zoster Virus (VZV) IgM**
Prod.No.:	**30114082**
Lot:	**VZVM-194**
Expiry:	**2027-08-31**
IFU:	**Rev. 01**

Component	Number	Lot	Prod. Date	Expiry Date
MTP	1	MM0490-194	2026-02	2027-08-31
CONJ	1	KM0490-194	2026-02	2027-08-31
CONTROL -	1	NM0490-194	2026-02	2027-08-31
CUT OFF	1	CM0490-194	2026-02	2027-08-31
CONTROL +	1	PM0490-194	2026-02	2027-08-31
WASH BUF 20x	1	W-1137	2026-02	2028-02-29
SOLN STOP	1	S-1069	2026-02	2028-02-29
SUB TMB	1	T-079	2025-12	2028-12-10
DIL M	1	DM-352	2026-03	2028-02-29

Component	OD Value	OD Validation Criteria
CONTROL -	0.020	< 0.200 < Cut-off
CUT OFF	0.523	0.150-1.300
CONTROL +	0.828	> Cut-off

Component	Target Value [NTU]	Target Range [NTU]
CONTROL +	19.06	10.48 - 27.63

The target value and target range of CONTROL + may be used optionally.

The positive control is valid if the absorbance value of CONTROL + > CUT OFF.

Final Report issued: 2026-03-30

Approved by: _____

Quality Control

Approved on: 2026-03-30

Gold Standard Diagnostics Frankfurt GmbH
Waldstrasse 23 A6
63128 Dietzenbach, Germany

IBL International GmbH
Flughafenstrasse 52a, 22335 Hamburg
Germany