

# Quality Control Certificate



## Reverse T3 (rT3) ELISA

REF 30123153


LOT 261530

2027-03-26  
YYYY-MM-DD

### Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.822	> 1.000	100%
CAL B	0.02	2.627	-	93%
CAL C	0.1	2.001	-	71%
CAL D	0.4	1.093	-	39%
CAL E	1	0.615	-	22%
CAL F	2	0.355	-	13%

	Result	Acceptable Range
OD Std. F/% BINDING	0.355/13%	< 20%



	Concentration found ng/mL	Target Value ng/mL	Acceptable range lower limit ng/mL	upper limit ng/mL
Control Low	CS3916	0.16	0.10	0.22
Control High	CS3916	0.45	0.33	0.55

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-05-07

**Accepted**

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)