

# Quality Control Certificate



## Adiponectin ELISA

REF 30126762


LOT 020226-2

2028-03-01  
YYYY-MM-DD

### Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	2	0.062	-	3%
CAL B	10	0.317	-	13%
CAL C	30	0.902	-	37%
CAL D	70	1.899	-	77%
CAL E	100	2.460	-	100%

	Target value ng/mL	Acceptable range ng/mL
CONTROL 1	6650	4660-8650
CONTROL 2	14860	10400-19320



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-05-05

# Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)