

Quality Control Certificate



Adiponectin ELISA

REF 30126762


LOT 180526-2

2028-06-17
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	2	0.071	-	2%
CAL B	10	0.344	-	12%
CAL C	30	1.016	-	35%
CAL D	70	2.236	-	77%
CAL E	100	2.901	-	100%

	Target value ng/mL	Acceptable range ng/mL
CONTROL 1	6930	4851-9009
CONTROL 2	15560	10892-20228



This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-09-11

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl