


Quality Control Certificate



Transglutaminase IgG ELISA

REF 30132293

LOT 24100241

 2026-05-31
YYYY-MM-DD

Final Release Results:

	Conc. U/mL	Measured value OD	Acceptable range		
			lower limit OD	target value OD	upper limit OD
Calibrator A	0	0.051			< 0.100
Calibrator F	100	2.328	> 1.000		
CONTROL -	-	0.200			≤ 0.300
CONTROL +	-	1.748	≥ 0.700		≤ 2.200

	Validation criteria (manual)			
	Measured value U/mL	Acceptable range		
		lower limit U/mL	target value U/mL	upper limit U/mL
CONTROL -	0.8			≤ 2.5
CONTROL +	26	17	27	37

Factor for qualitative assay evaluation	0.31
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2025-01-13

This document has been produced electronically and is valid without a signature.