


Quality Control Certificate



Transglutaminase IgA ELISA

REF 30132295

LOT 25070133

 2026-12-31
YYYY-MM-DD

Final Release Results:

	Conc.	Measured value	Acceptable range		
	U/mL		lower limit	target value	upper limit
		OD	OD	OD	OD
Calibrator A	0	0.052			< 0.100
Calibrator F	100	2.456	> 1.000		
CONTROL -	-	0.204			≤ 0.300
CONTROL +	-	1.272	≥ 0.500		≤ 1.600

	Validation criteria (manual)			
	Measured value	Acceptable range		
		lower limit	target value	upper limit
U/mL	U/mL	U/mL	U/mL	
CONTROL -	2.2			≤ 2.5
CONTROL +	28	16	25	34

Factor for qualitative assay evaluation	0.21
---	------

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2025-07-28

This document has been produced electronically and is valid without a signature.