


Quality Control Certificate



sCD40L ELISA

REF 30150492

LOT 309711-032

 2027-07-31
YYYY-MM-DD

| | Acceptable range | | |
|--------------|----------------------|-----------------|----------------------|
| | lower limit ng/mL | target ng/mL | upper limit ng/mL |
| CONTROL HIGH | 20 | 35 | 50 |
| CONTROL LOW | 1 | 3 | 5 |

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Accepted

Validation date: 2024-10-24

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl