


# Quality Control Certificate



## sCD40L ELISA

REF 30150492

LOT 401155-013

 2028-07-31  
YYYY-MM-DD

	Acceptable range		
	lower limit ng/mL	target ng/mL	upper limit ng/mL
CONTROL HIGH	20	35	50
CONTROL LOW	1	3	5

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

**Accepted**

Validation date: 2025-09-12

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)