

Quality Control Certificate



HMGB1 Express ELISA

REF 30164033


LOT EHMx135

2027-01-31
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL 1	0	0.049	-	3%
CAL 2	2.5	0.073	-	4%
CAL 3	5.0	0.124	-	7%
CAL 4	10	0.229	-	13%
CAL 5	20	0.430	-	24%
CAL 6	40	0.880	-	50%
CAL 7	80	1.766	-	100%

	Concentration found ng/mL	Acceptable range		
		lower limit ng/mL	target value ng/mL	upper limit ng/mL
CONTROL +	21	14	21	27



Kit Composition:

REF	LOT	Icon	REF	LOT	Icon
30164035 MTP	K402	🕒	30164042 ENZCONJDIL	K407	🕒
30164041 LYO CONC ENZCONJ	K403	🕒	30170004 CONC WASHBUF	K415	🕒
30164037 CAL	K404	🕒	30169998 COLREA A	K297	🕒
30164038 CONTROL +	K405	🕒	30170002 COLREA B	K298	🕒
30164040 DILBUF	K406	🕒	30170003 COLOUR STOP	K300	🕒

Valid edition of the instructions for use: 2023-07

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2025-11-18

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl