

Quality Control Certificate



HMGB1 Express ELISA

REF 30164033


LOT EHMx136

2027-07-31
YYYY-MM-DD

Final Release Results:

	Concentration	Measured value	Range	OD/ODmax
	ng/mL	OD	OD	
CAL 1	0	0.033	-	2%
CAL 2	2.5	0.083	-	5%
CAL 3	5.0	0.117	-	7%
CAL 4	10	0.231	-	14%
CAL 5	20	0.446	-	27%
CAL 6	40	0.855	-	52%
CAL 7	80	1.643	-	100%

	Concentration found	Acceptable range		
	ng/mL	lower limit	target value	upper limit
	ng/mL	ng/mL	ng/mL	ng/mL
CONTROL +	19	15	22	28



Kit Composition:

REF	LOT	
30164035 MTP	K712	2027-07
30164041 LYO CONC ENZCONJ	K713	2027-07
30164037 CAL	K714	2027-07
30164038 CONTROL +	K715	2027-07
30164040 DILBUF	K716	2027-07

REF	LOT	
30164042 ENZCONJDIL	K717	2027-07
30170004 CONC WASHBUF	K711	2027-09
30169998 COLREA A	K708	2027-09
30170002 COLREA B	K709	2027-09
30170003 COLOUR STOP	K710	2028-03

Valid edition of the instructions for use: 2023-07

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-06-24

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl