

# Quality Control Certificate



## Acetylcholine Receptor Autoantibodies (ARAb) RRA


<b>REF</b> 30221148 & 30221149	<b>LOT</b> RAR565	<b>2026-06-09</b> YYYY-MM-DD
--------------------------------	-------------------	---------------------------------

### Final Release Results:

		Measured value	
Total Activity	TA	83491	cpm
Binding of Standard A	B0/TA	1.5	%
Maximum Binding	Bmax/TA	24	%

	Concentration	Measured value	Range	B/Bmax
	nmol/L	cpm	cpm	
CAL A	0	1262	-	0%
CAL B	0.2	2718	-	7%
CAL C	0.5	5122	-	20%
CAL D	1.2	10661	-	48%
CAL E	3	16638	-	78%
CAL F	8	20938	-	100%

	Concentration found	Acceptable range		
	nmol/L	lower limit	target value	upper limit
	nmol/L	nmol/L	nmol/L	nmol/L
CONTROL CO	0.42	0.34	0.43	0.51
CONTROL +	1.16	0.81	1.16	1.51



### Kit Composition:

REF		LOT	
KRAR013	LYO TRACER	H672	2026-06-09
KRAR031/33	ANTISERUM	H571-D	2026-06
KRAR111A	CAL A	H450-2A	2026-06-09
KRAR111B-F	CAL B-F	H440-8G	2026-06-09
KRAR151	CONTROL CO	H441-8G	2026-06-09

REF		LOT	
KRAR161	CONTROL +	H442-8G	2026-06-09
KRAR253	WASHBUF	H616	2028-01
KRAR261	BUF	H542-E	2026-06

Valid edition of the instructions for use: 2023-08

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-03-17

# Accepted

This document has been produced electronically and is valid without a signature.



**IBL International GmbH**  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)