

Quality Control Certificate



Acetylcholine Receptor Autoantibodies (ARAb) RRA


REF 30221148 & 30221149	LOT RAR567	2026-08-04 YYYY-MM-DD
--------------------------------	-------------------	---------------------------------

Final Release Results:

		Measured value	
Total Activity	TA	86079	cpm
Binding of Standard A	B0/TA	1.3	%
Maximum Binding	Bmax/TA	18	%

	Concentration nmol/L	Measured value cpm	Range cpm	B/Bmax
CAL A	0	1098	-	0%
CAL B	0.2	2324	-	8%
CAL C	0.5	4363	-	21%
CAL D	1.2	8993	-	50%
CAL E	3	13524	-	79%
CAL F	8	16810	-	100%

	Concentration found nmol/L	Acceptable range		
		lower limit nmol/L	target value nmol/L	upper limit nmol/L
CONTROL CO	0.42	0.34	0.43	0.51
CONTROL +	1.17	0.81	1.16	1.51



Kit Composition:

REF		LOT	
KRAR013	LYO TRACER	H707	2026-08-04
KRAR031/33	ANTISERUM	H571-F	2026-08
KRAR111A	CAL A	H450-4C	2026-08-04
KRAR111B-F	CAL B-F	H440-10I	2026-08-04
KRAR151	CONTROL CO	H441-10I	2026-08-04

REF		LOT	
KRAR161	CONTROL +	H442-10I	2026-08-04
KRAR253	WASHBUF	H683	2028-03
KRAR261	BUF	H661-B	2026-08

Valid edition of the instructions for use: 2023-08

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-05-13

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl