

Quality Control Certificate




Cortisol Saliva Luminescence Immunoassay

REF 30221150 & 30221151	LOT LCL207	2027-01-31 YYYY-MM-DD
--------------------------------	-------------------	---------------------------------

Final Release Results:

	Concentration µg/dL	Measured value RLU	Range RLU	RLU/RLUmax
CAL A	0	49377513	-	100%
CAL B	0.015	36431848	-	74%
CAL C	0.03	29652063	-	60%
CAL D	0.10	15594808	-	32%
CAL E	0.40	5049580	-	10%
CAL F	1.00	2071250	-	4%
CAL G	3.20	581053	-	1%

	Concentration found µg/dL	Acceptable range		
		lower limit µg/dL	target µg/dL	upper limit µg/dL
CONTROL 1	0.083	0.051	0.078	0.106
CONTROL 2	0.72	0.47	0.67	0.88



Kit Composition:

REF	LOT		REF	LOT	
KLCL651/9	CONTROL 1	L488B-3A	2027-07	KLZZ641/9	LUMINREAG 1
KLCL661/9	CONTROL 2	L489B-3A	2027-07	KLZZ651/9	LUMINREAG 2
KLCL511	MTP	K113	2027-05		
KLCL541/9	ENZCONJ	K114	2027-08		
KLCL611/9	CAL A	L486-3A	2027-07		
KLCL621/9	CAL B-G	L487B-3A	2027-07		
KLZZ709	CONC WASHBUF	L970	2027-05		

Valid edition of the instructions for use: 2024-11

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-08-01

Accepted

This document has been produced electronically and is valid without a signature.