

Quality Control Certificate




Metanephrine ELISA

REF 30226921	LOT EMN206U	2027-04-30 YYYY-MM-DD
---------------------	--------------------	---------------------------------

Final Release Results:

	Concentration µg/L	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.555	> 1.000	100%
CAL B	26	2.431	-	95%
CAL C	64	2.114	-	83%
CAL D	160	1.470	-	58%
CAL E	400	0.959	-	38%
CAL F	1000	0.621	-	24%
CAL G	2500	0.332	< 0.500	13%

	Concentration found µg/L	Acceptable range		
		lower limit µg/L	target µg/L	upper limit µg/L
CONTROL 1	124.5	75.5	117.9	160.4
CONTROL 2	388	312	487	663



Kit Composition:

REF		LOT	
KEZZ621	CONTROL 1	L945	2027-04
KEZZ631	CONTROL 2	L946	2027-04
KEZZ811	MTP	L535	2027-08
KEZZ661	HYDRUB	L948	2030-04
KEZZ671	HCL	L949	2028-04
KEZZ651	INDICATORBUF	L792	2027-12
KEZZ641	ACYLREAG	L947	2027-04
KEMN541	CONC ENZCONJ	L943	2027-04

REF		LOT	
KEZZ611	CAL A-G	L944	2027-04
KEMN533	BIOTIN	L474-A	2027-04
KEMN521	ANTISERUM	L942	2027-04
KEZZ681	CONC ASSAYBUF	L950	2027-04
KLZZ879	CONC WASHBUF	L951	2027-04
KEZZ961	PNPP SUBS	K180	2027-09
KEZZ951	PNPP STOP	L952	2028-04

Valid edition of the instructions for use: 2024-04

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-07-31

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl