

Quality Control Certificate



Steroid Tuning Mix 1

REF 30227628	LOT MTSA113	2026-06-30 YYYY-MM-DD
---------------------	--------------------	---------------------------------

Final Release Results:

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

All results were acquired using this mass transitions in positive electro spray ionization mode (ESI):

Analyte	Q1	Q3	Internal Standard	Q1	Q3
dihydrotestosterone	291	255	dihydrotestosterone-D3	294	258
testosterone	289	97	testosterone-13C3	292	100
21-deoxycortisol	347	311	21-deoxycortisol-D8	355	319
17-hydroxyprogesterone	331	97	17-hydroxyprogesterone-D8	339	100
androstenedione	287	97	androstenedione-13C3	290	100

Kit Composition:

REF	LOT	
30227628 TUNING 1	L765	2026-06

Valid edition of the instructions for use: 2023-08

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: M.Dries

Validation date: 2025-01-02

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl