

Quality Control Certificate



Steroid Tuning Mix 2

REF 30227629	LOT MTSB113	2026-05-31 YYYY-MM-DD
---------------------	--------------------	---------------------------------

Final Release Results:

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

All results were acquired using this mass transitions in positive electro spray ionization mode (ESI):

Analyte	Q1	Q3	Internal Standard	Q1	Q3
11-deoxycorticosterone	331	97	11-deoxycorticosterone-13C3	334	100
11-deoxycortisol	347	97	11-deoxycortisol-D5	352	100
dehydroepiandrosterone	289	271	dehydroepiandrosterone-D5	294	258
aldosterone*	359	189	aldosterone-D4*	363	190
dehydroepiandrosterone-sulfate	271	253	dehydroepiandrosterone-sulfate-D5	276	258

*negative ionization mode

Kit Composition:

REF	LOT	
30227629 TUNING 2	L766	2026-05

Valid edition of the instructions for use: 2023-08

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: M.Dries

Validation date: 2024-12-04

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl