

Quality Control Certificate



Steroid Tuning Mix 2

REF 30227629

LOT MTSB114



2029-01-31

YYYY-MM-DD

Final Release Results:

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

All results were acquired using this mass transitions in positive electro spray ionization mode (ESI):

Analyte	Q1	Q3	Internal Standard	Q1	Q3
11-deoxycorticosterone	331	97	11-deoxycorticosterone-13C3	334	100
11-deoxycortisol	347	97	11-deoxycortisol-D5	352	100
dehydroepiandrosterone	289	271	dehydroepiandrosterone-D5	294	258
aldosterone*	359	189	aldosterone-D4*	363	190
dehydroepiandrosterone-sulfate	271	253	dehydroepiandrosterone-sulfate-D5	276	258

*negative ionization mode

Kit Composition:

REF

LOT



30227501 TUNING 2

K990

2029-01

Valid edition of the instructions for use: 2023-08

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-10-06

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl