

Quality Control Certificate



Steroid Tuning Mix 3

REF 30227630

LOT MTSC113A

2027-03-31
YYYY-MM-DD

Final Release Results:

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

All results were acquired using this mass transitions in positive electro spray ionization mode (ESI):

Analyte	Q1	Q3	Internal Standard	Q1	Q3
cortisol	363	121	cortisol-D4	367	121
cortisone	361	163	cortisone-13C3	364	166
dexamethasone	393	373	dexamethasone-D5	398	378
melatonin	233	174	melatonin-D4	237	178
corticosterone	347	121	corticosterone-13C3	350	332

Kit Composition:

REF	LOT	
30227502 TUNING 3	L767A	2027-03

Valid edition of the instructions for use: 2023-08

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-07-02

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl