

Quality Control Certificate



Steroid Tuning Mix 4

REF 30227631

LOT MTSD113A



2027-03-31

YYYY-MM-DD

Final Release Results:

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

All results were acquired using this mass transitions in negative electro spray ionization mode (ESI):

Analyte	Q1	Q3	Internal Standard	Q1	Q3
estradiol	271	145	estradiol-13C3	274	148
estrone	269	145	estrone-13C3	272	148
estriol	287	171	estriol-13C3	290	174
progesterone*	315	97	progesterone-D9*	324	100
17-hydroxypregnenolone	331	287	17-hydroxypregnenolone-13C2-D2	335	316

*positive ionization mode

Kit Composition:

REF

LOT



30227503 TUNING 4

L772A

2027-03

Valid edition of the instructions for use: 2023-08

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: T. Borchert

Validation date: 2026-07-02

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl