

QUALITY CONTROL CERTIFICATE

Glucagon ELISA

REF CAN-GLN-390

LOT 261220

 **2027-02-05**

YYYY-MM-DD

Standard Curve

Calibrator	Concentration (pg/mL)	Mean OD (450 nm)	% Binding (OD/OD _{MAX})
A	0	0.063	2 %
B	11	0.141	6 %
C	22	0.230	9 %
D	44	0.416	17 %
E	88	0.754	30 %
F	176	1.324	53 %
G	352	2.517	100 %

Parameter	Result	Acceptable Range
Calibrator A - OD (450 nm)	0.063	< 0.120
Calibrator G - OD (450 nm)	2.517	> 0.600

Kit Controls

Control			Result (pg/mL)	Target Value (pg/mL)	Acceptable Range (pg/mL)	
					Low	High
1	LOT	CS3883	36.3	35.9	26.9	44.9
2	LOT	CS3883	211.1	209.0	156.8	261.3

The above product has met all quality specifications, been reviewed by a quality control representative and released for sale.

This document has been produced electronically and is valid without a signature.

Date: 2026-04-09

QC Representative: Yegor Zarubin

YYYY-MM-DD