

Quality Control Certificate



Interleukin-10 high sensitivity ELISA

REF BE58101

LOT 490667-003

 2029-02-28
YYYY-MM-DD

	Acceptable range		
	lower limit pg/mL	target pg/mL	upper limit pg/mL
Positive Control	10	55	100

	Dilution
Amplifications Reagent 1	1:100
Amplifications Reagent 2	1:100

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-08-26

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl