

Quality Control Certificate



Angiotensin I (PRA) ELISA

REF DB52011


LOT 260820

2027-03-04
YYYY-MM-DD

Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.431	> 1.200	100%
CAL B	0.2	2.194	-	90%
CAL C	0.5	1.935	-	80%
CAL D	1.5	1.331	-	55%
CAL E	4	0.773	-	32%
CAL F	10	0.412	-	17%
CAL G	25	0.225	-	9%
CAL H	60	0.141	-	6%

	Result	Acceptable Range
OD Std. H/% BINDING	0.141/6%	< 10%



	Concentration found ng/mL	Target Value ng/mL	Acceptable range lower limit ng/mL	upper limit ng/mL
KIT CONTROL Low CS3867	0.80	0.82	0.62	1.03
KIT CONTROL High CS3867	9.8	9.8	7.4	12.3

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-04-08

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl