

Quality Control Certificate




5alpha-Dihydrotestosterone ELISA

| | | | | | |
|------------|----------------|------------|---------------|--|-------------------|
| REF | DB52021 | LOT | 260570 | | 2027-02-03 |
| | | | | | YYYY-MM-DD |

Final Release Results:

| | Concentration pg/mL | Measured value OD | Range OD | OD/ODmax |
|-------|------------------------|-------------------------|-------------|----------|
| CAL A | 0 | 2.735 | >0.600 | 100% |
| CAL B | 25 | 2.296 | - | 84% |
| CAL C | 100 | 1.755 | - | 64% |
| CAL D | 250 | 1.311 | - | 48% |
| CAL E | 500 | 0.932 | - | 34% |
| CAL F | 1000 | 0.651 | - | 24% |
| CAL G | 2500 | 0.398 | - | 15% |



| | Result | Acceptable Range |
|----------------------------|-----------|------------------|
| CAL G % BINDING (OD/ODmax) | 0.398/15% | <25 % |

| | Concentration found pg/mL | Target Value pg/mL | Acceptable range lower limit pg/mL | upper limit pg/mL |
|-------------------------|---------------------------------|--------------------------|--|----------------------|
| KIT CONTROL High CS3838 | 438.4 | 430.6 | 323.0 | 538.3 |
| KIT CONTROL Low CS3838 | 129.5 | 128.7 | 77.2 | 180.2 |

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-03-03

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl