

Quality Control Certificate




5alpha-Dihydrotestosterone ELISA

REF	DB52021	LOT	261520		2027-04-11
					YYYY-MM-DD

Final Release Results:

	Concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.590	>0.600	100%
CAL B	25	2.199	-	85%
CAL C	100	1.618	-	62%
CAL D	250	1.202	-	46%
CAL E	500	0.881	-	34%
CAL F	1000	0.603	-	23%
CAL G	2500	0.344	-	13%



	Result	Acceptable Range
CAL G % BINDING (OD/ODmax)	0.344/13%	<25 %

	Concentration found pg/mL	Target Value pg/mL	Acceptable range lower limit pg/mL	upper limit pg/mL
KIT CONTROL High CS3920	416.2	395.7	296.8	494.6
KIT CONTROL Low CS3920	127.8	129.2	77.5	180.9

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-05-07

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl