

Quality Control Certificate




5alpha-Dihydrotestosterone ELISA

| | | | | | |
|------------|----------------|------------|---------------|--|-------------------|
| REF | DB52021 | LOT | 261970 | | 2027-05-27 |
| | | | | | YYYY-MM-DD |

Final Release Results:

| | Concentration pg/mL | Measured value OD | Range OD | OD/ODmax |
|-------|------------------------|-------------------------|-------------|----------|
| CAL A | 0 | 2.469 | >0.600 | 100% |
| CAL B | 25 | 2.075 | - | 84% |
| CAL C | 100 | 1.542 | - | 62% |
| CAL D | 250 | 1.157 | - | 47% |
| CAL E | 500 | 0.820 | - | 33% |
| CAL F | 1000 | 0.570 | - | 23% |
| CAL G | 2500 | 0.329 | - | 13% |



| | Result | Acceptable Range |
|----------------------------|-----------|------------------|
| CAL G % BINDING (OD/ODmax) | 0.329/13% | <25 % |

| | Concentration found pg/mL | Target Value pg/mL | Acceptable range lower limit pg/mL | upper limit pg/mL |
|-------------------------|---------------------------------|--------------------------|--|----------------------|
| KIT CONTROL High CS3961 | 392.9 | 398.2 | 298.7 | 497.8 |
| KIT CONTROL Low CS3961 | 114.4 | 112.6 | 67.6 | 157.6 |

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-06-18

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl