

# Quality Control Certificate



## Pregnenolone ELISA

REF DB52031


LOT 254300

2027-05-31  
YYYY-MM-DD

### Final Release Results:

	Concentration ng/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.158	> 1.500	100%
CAL B	0.1	1.542	-	71%
CAL C	0.4	1.080	-	50%
CAL D	1.6	0.659	-	31%
CAL E	6.4	0.313	-	15%
CAL F	25.6	0.135	-	6%

	Result	Acceptable Range
OD Std. F/% BINDING	0.135/6%	< 30%



	Concentration found ng/mL	Target Value ng/mL	Acceptable range lower limit ng/mL	upper limit ng/mL
Control Low	CS3788	1.26	0.94	1.56
Control High	CS3788	5.53	3.68	6.14

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2026-02-12

**Accepted**

This document has been produced electronically and is valid without a signature.



IBL International GmbH  
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11  
E-MAIL: [ibl@tecan.com](mailto:ibl@tecan.com)  
WEB: [www.tecan.com/ibl](http://www.tecan.com/ibl)