

Quality Control Certificate



Estrone ELISA


REF DB52051

LOT 251300

2026-10-08
YYYY-MM-DD

Final Release Results:

	Concentration pg/mL	Measured value OD	Range OD	OD/ODmax
CAL A	0	2.786	> 1.1	100%
CAL B	20	2.423	-	87%
CAL C	60	1.967	-	71%
CAL D	200	1.388	-	50%
CAL E	600	0.825	-	30%
CAL F	2000	0.427	-	15%



	Result	Acceptable Range
OD Std. F/% BINDING	0.427/15%	< 20%

	Concentration found pg/mL	Target Value pg/mL	Acceptable range lower limit pg/mL	upper limit pg/mL
KIT CONTROL Low CS3546	83.0	82.5	49.5	115.5
KIT CONTROL High CS3546	514.4	504.9	378.7	631.1

This product has been tested successfully by the Quality Control Department and was released for sale according to the existing specifications.

Verified by: Dr. Spangenberg

Validation date: 2025-05-22

Accepted

This document has been produced electronically and is valid without a signature.



IBL International GmbH
Flughafenstr. 52A, 22335 Hamburg, Germany

Tel.: + 49 (0) 40 532891 -0 Fax: -11
E-MAIL: ibl@tecan.com
WEB: www.tecan.com/ibl